

Certificate of Quality

The Membrane Prefilter Capsule has been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

ClariPro GK Membrane Prefilter Capsule

Catalog No. : LGKX5404EEXX101
 Type : LGK
 Pore Size : 0.5 µm (Nominal)
 Lot Number : LK8817G SI.No. 002

SPECIFICATION

Length	10"
Filter Media	Polyethersulfone Membrane with Microglassfiber Prefilter
Drainage Layers	Polyester
Housing	Polypropylene
Differential Pressure	< 4Kg/cm ² at 30 °C
Maximum Operating Temperature	80 °C @ < 2 Kg/cm ²

LOT RELEASE CRITERIA

100% Integrity Tested : The capsule filter has been tested for integrity by Bubble point Test using 50% IPA/Water solution. Bubble point was: ≥ 5 psi (0.33 Bar)

Water Flow Rate : ≥ 15 lpm @ 0.70 Kg/cm² @ 27 °C

VALIDATED FOR

Sterilization : Maintains integrity after 5 autoclaving cycles at 125 °C of 30 minutes each.

Extractable : Within limits as specified in USP.

Oxidizable matter : Passes test as per USP.

Biosafety : Passes Biological Reactivity Tests, *In Vivo* for Class VI plastic as described in USP <88>.

Indirect Food Additives : Passes as per FDA 21CFR 177.1520(a)1(i).

Particle Release : Passes test as per USP <788>, "Particulate matter in Injections".

Fiber Release : Complies with FDA 21CFR 210.3(b)(6).

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.



Head of Quality Assurance

Issue Date: 24-Jul-17

Advanced Microdevices Pvt. Ltd.
 21, Industrial Area, Ambala Cantt, INDIA,
 Tel: +91-171-2699290/ 2699274
 Website: www.mdimembrane.com
 Email: info@mdimembrane.com

An ISO 9001 Company